

Republic of the Philippines BULACAN STATE UNIVERSITY City of Malolos

## BIDS AND AWARDS COMMITTEE FOR GOODS, SUPPLIES, AND SERVICES

December 1, 2023

### **Bid Bulletin No. 1**

### Modifications in the Invitation to Bid, Bid Data Sheet, Schedule of Requirements, and Technical Specifications

This Bid Bulletin no. 1 is issued to modify or amend items in the Bidding Documents for project: "Supply and Delivery of Nursing Maternal and Neonatal Simulation System (G-2023-54)". This shall form an integral part of the said Bidding Documents.

### I. SECTION I. INVITATION TO BID

	FROM	ТО
2	The Bulacan State University (BulSU) now invites bids for the above Procurement Project. Delivery of the goods is required within thirty (30) calendar days upon receipt of Notice to Proceed. Bidders should have completed, within the last three (3) years from the date of submission and receipt of bids on December 12, 2023, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).	The Bulacan State University (BulSU) now invites bids for the above Procurement Project. Delivery of the goods is required <i>within ninety (90) calendar days</i> <i>upon receipt of Notice to Proceed</i> . Bidders should have completed, within the last three (3) years from the date of submission and receipt of bids on December 12, 2023, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).

### II. SECTION III. BID DATA SHEET

	FROM	ТО			
5.3	For this purpose, contracts similar to the	For this purpose, contracts similar to the			
	<ul> <li>Project shall be:</li> <li>a. Similar Project: Research/ Science Laboratory</li> <li>b. Project equivalent to at least fifty percent (50%) of the ABC amounting to ₱ 575,000.00</li> <li>c. Completed within the last three (3 years prior to the deadline for the submission and receipt of bids.</li> </ul>	Laboratory <u>Equipment</u> bProject equivalent to at least fifty percent (50%) of the ABC amounting to ₱ 575,000.00 c. Completed within the last three (3)			

### **III. SECTION VI. SCHEDULE OF REQUIREMENTS**

FROM	ТО			
The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.	The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.			
Within thirty (30) calendar days upon receipt of Notice to Proceed	Within ninety (90) calendar days upon receipt of Notice to Proceed.			

### IV. SECTION VII. TECHNICAL SPECIFICATIONS

FROM	ТО			
	<u>Please see attached Annex "A" for the</u> <u>revised Technical Specifications.</u>			

This Supplemental/ Bid Bulletin no. 1 is issued this  $1^{st}$  day of December, 2023 for guidance and information of all concerned.

For information and guidance of all concerned.

# Prof. YOLANDA C. ROBERTO

Chairperson BAC for Goods, Supplies & Services

## Section VII. Technical Specifications

### **Instruction to Bidders:**

Please fill up the form accordingly, state either "**COMPLY**" or "**NOT COMPLY**" in each of the individual parameters of each specification, and sign the conforme below. Failure to conform will result in a rating of "**FAILED**".

Item	Description	Qty./ unit	Statement of Compliance		
1	Maternal and Neonatal birthing simulation system	1 pc	[Bidders must state here		
	Automatic birthing system,		either <u>"Comply" or "Not</u>		
	Full-body newborn airway trainer,		<u>Comply</u> against each of the		
	Articulating birthing fetus,		individual parameters of each		
	Two umbilical cords,		Specification stating the corresponding performance		
	Two dilating cervices,		parameter of the equipment		
	Two umbilical clamps,		offered. Statements of		
	Two vulval inserts		"Comply" or "Not Comply"		
	Three vulvae for postpartum suturing,		must be supported by		
	International power supply 100-240 VAC		evidence in a Bidders Bid and		
	CPR Link software.		cross-referenced to that		
	CPR Link connection cable,		evidence. Evidence shall be		
	CPR module preinstalled in manikin		in the form of manufacturer's		
	Silicone lubricant,		un-amended sales literature, unconditional statements of		
	with Carrying bag		specification and compliance		
2	Newborn Simulator (Female)	3 pcs	issued by the manufacturer,		
	Very useful for pre-birth courses.		samples, independent test		
	This true-to-life model of a newborn male		data etc., as appropriate. A		
	baby is suitable for a variety of different		statement that is not		
	training procedures: Changing the baby's		supported by evidence or is		
	nappy, Feeding and administration of		subsequently found to be		
	medicine Obtaining urine specimens,		contradicted by the evidence presented will render the Bid		
	Dressing and undressing, Picking up,		under evaluation liable for		
	carrying, and positioning of a baby (head tilts		rejection. A statement either		
	backwards if the baby is incorrectly held),		in the Bidder's statement of		
	Washing, cleaning of mouth, eyes, ear and		compliance or the supporting		
	nose, a rectal temperature, Examination of fontanelles and testicles		evidence that is found to be false either during Bid		
	Head Circumference,		evaluation, post-qualification		
	Approx. 48 cm		or the execution of the		
	Approx: 34 cm;		Contract may be regarded as		

	Weight 3.0 kgs		fraudulent and render the			
3	Newborn Simulator (Male)	2 pcs	Bidder or supplier liable for			
	Very useful for pre-birth courses.		prosecution subject to the			
	This true-to-life model of a newborn male		applicable laws and			
	baby is suitable for a variety of different		issuances.] [Bidders must			
	training procedures: Changing the baby's		state here either "Comply" or "Not Comply" against each of			
	nappy, Feeding and administration of		the individual parameters of			
	medicine Obtaining urine specimens,		each Specification stating the			
	Dressing and undressing, Picking up,		corresponding performance			
	carrying, and positioning of a baby (head tilts		parameter of the equipment			
	backwards if the baby is incorrectly held),		offered. Statements of			
	Washing, cleaning of mouth, eyes, ear and		"Comply" or "Not Comply"			
	nose, a rectal temperature, Examination of fontanelles and testicles		must be supported by evidence in a Bidders Bid and			
	Head Circumference,		cross-referenced to that			
	Approx. 48 cm		evidence. Evidence shall be			
	Approx: 34 cm;		in the form of manufacturer's un-amended sales literature, unconditional statements of			
	Weight 3.0 kgs					
			specification and compliance			
	Warranty		issued by the manufacturer,			
	Minimum of 1-year warranty		samples, independent test			
			data etc., as appropriate. ]			
	COMPLIANCE					
	1. With Brochure / catalogue of the item being offered					
	2. With Manufacturer's Authorization Form or Certificate as Authorized Distributor/Reseller or Certificate of Registration as Manufacturer/Fabricator					
	3. With Warranty Certificate from Manufacturer and/or Supplier/Authorized Reseller					
	Contract Implementation:					
	<b>Contract Implementation:</b> The supplier must advise the Supply Office and End- user unit of the delivery schedule at least three (3) calendar days before the actual delivery.					

Prepared by:

### Mr. MARK RONALD DE LEON

Technical Working Group

### Mr. MARK ANTHONY LODRIGITO

Technical Working Group

I understand and swear to comply and deliver all the above requirements.

### **CONFORME:**

Company Name	: .					
Address	: .					
Complete Name of						
Authorized Representative:						
Signature	:					
Date	: .				 	