



REOUEST FOR OUOTATION FOR THE PROCUREMENT OF GOODS AND SERVICES

| | |
|---------------------------------|---|
| **Mandatory to fill in** | |
| COMPANY NAME: | Quotation No. 20-271-10 |
| ADDRESS: | Purchase Request No. F-2020-10-0165 |
| CONTACT No. | Purpose: Emergency medical equipment as a preventive measure against COVID-19 |
| TIN No. | ABC: 649,000.00 |
| PhilGEPS Registration No. | Delivery Period: 7 Calendar Days upon receipt of Purchase Order |
| EMAIL ADDRESS: | |

INSTRUCTIONS TO SUPPLIERS:

1. Please quote your lowest price on the item/s listed below comprising the necessary taxes.

2. It is mandatory to indicate the brand and/or model of the items being offered and to attach a brochure thereof whenever applicable.

3. Indicate the warranty period in cases of equipment or whenever applicable.

4. Forthwith submit the accomplished quotation duly signed by your representative.

5. Suppliers are required and mandated to attach and submit the following documentary requirements:
a) Valid Mayor's/ Business Permit; b) BIR Certificate of Registration; c) Authority to Print Receipt; d) DTI/SEC Registration; e) PhilGEPS Membership Certificate; and f) PCAB License in cases of Infrastructure projects.

6. All items must conform with the internationally accepted standard and sub-standard items shall not be accepted.

| ITEM NO. | ITEM & SPECIFICATION | QTY/UNIT | BRAND & MODEL OFFERED | UNIT PRICE | TOTAL PRICE |
|----------|---|----------|-----------------------|------------|-------------|
| 1 | Automatic Alcohol Spray Dispenser with Stand | 65 pcs | | | |
| 2 | Handsfree Non-Contact Forehead Body Infrared Thermal Scanner and Automatic Spray Dispenser with Stand | 20 units | | | |
| | SPECIFICATIONS: 1. Measurement method: No contact with forehead of infrared thermometer 2. Theory: F and C degrees (Fahrenheit and Certigrade) 3. Accuracy tolerance: +/- 0.2 (34 ~ 45 degree C) 4. Automatic alarm for abnormal temperature: Red light flashes, continuous alarm sound "Di,Di..." 5. Distance: 5 ~ 10 cm 6. Standby: 5s 7. Placement: Hanging / double-sided adhesive tape / bracket fixed 8. Display: Digital display 9. Measuring temperature: 0 ~ 80 degree C 10. Ambient temperature: 5 ~ 45 degree C 11. Response time: 500 ms 12. Working wave: 5 microns 13. Automaitc Alcohol Spray Dispenser 14. Stand/Tripod | | | | |
| 3 | Face Shield | 700 pcs | | | |
| 4 | Powered Air Purifying Respirator with Hood | 2 units | | | |
| | Specifications: -CE Certified -Adjustable 3 level air flow control -Rechargeable battery 4-8 usage -99.996% Filter efficiency | | | | |
| 5 | Ultraviolet Rays Sterilization Implement | 4 units | | | |
| | Specifications: -Built-in Timer -Double-lamp. 30Wx2pcs -Replaceable lamp -Wavelength:250- 270 nanometer | | | | |
| | | | | | |
| | NOTE: | | | | |
| | 1. WITH WARRANTY | | | | |
| | 2. CE CERTIFIED | | | | |

| | |
|---|--|
| Accomplished by: | By the authority of the University President. |
| <div>Supplier's Representative (Print name and Signature)</div> | <div>Assoc. Prof. JOSEPH ROYE CELESTINO BAC Chairman</div> |
| Date Accomplished : _____ | Canvassed by: |
| | <div>Name and Signature</div> |
| BulSU-OP-PU-03F3 Revision: 1 JAM/10-22-20 | |