



Republic of the Philippines
BULACAN STATE UNIVERSITY
City of Malolos

BIDS AND AWARDS COMMITTEE FOR GOODS AND SERVICES

7 December 2020

Bid Bulletin No. 1
Modification in the Technical Specifications

This Bid Bulletin no. 1 is being issued to modify or amend items in the Bidding Documents for the project – **Supply and Delivery of LIFE SCIENCE and CAESAR Center Indoor Laboratory Equipment (BulSU-BAC-G-2020-33)**. This shall form an integral part of the said Bidding Documents.

TECHNICAL SPECIFICATIONS

FROM	TO
ANNEX 10: REFRIGERATED CENTRIFUGE (2 sets) Drive Mechanism: Brushless Induction	ANNEX 10: REFRIGERATED CENTRIFUGE (2 sets) Drive Mechanism: Brushless Induction or Equivalent
Speed Setting Range: 300 - 18,000 rpm (100 rpm increments)	Speed Setting Range: 300 - 18,000 rpm (10 rpm increments)
Speed Accuracy: ± 20 rpm	This requirement was removed.
Power Supply: Single Phase AC220V, 50 Hz , 10A	Power Supply: Single Phase AC220V, 50 Hz , 10A (Supplier must provide a step down or frequency converter)
New and/or additional specification	2 sets Centrifuge tube (1 set for Micro tube angle rotor; 1 set for Micro tube Fixed rotor)
New and/or additional specification	Supplier must have a partner Lab that has licensed chemist that can provide extensive application support.

This Supplemental/ Bid Bulletin no. 1 is issued this 7th day of December 2020 for guidance and information of all concerned.


Assoc. Prof. JOSEPH ROY E. CELESTINO
Chairperson
Bids and Awards Committee for Goods & Services