



REQUEST FOR QUOTATION FOR THE PROCUREMENT OF GOODS AND SERVICES

****Mandatory to fill in****

COMPANY NAME:	Quotation No. 22-11-792
CONTACT No.	Purchase Request No. F-2022-11-0534
Address:	Purpose: For laboratory use
TIN No.	ABC: 64,000.00
PhilGEPS Registration No.	Delivery Period: 7 Calendar Days upon receipt of Purchase Order
EMAIL ADDRESS:	

INSTRUCTIONS TO SUPPLIERS:

- Please **quote your lowest price** on the item/s listed below **comprising the necessary taxes**.
- It is mandatory to **indicate the brand** and/**or model** of the items being offered and to **attach a brochure** thereof whenever applicable
- Indicate the **warranty period** in cases of equipment or whenever applicable.
- Forthwith submit the accomplished quotation **duly signed by your representative**.
- Suppliers are required and mandated to attach and submit the following documentary requirements:
a) Valid Mayor's/ Business Permit; b) BIR Certificate of Registration; c) Authority to Print Receipt; d) PhilGEPS Membership Certificate and e) Omnibus Sworn Statement
- All items must conform with the **internationally accepted standard** and **sub-standard items shall not be accepted**.

ITEM NO.	ITEM & SPECIFICATION	QTY/UNIT	BRAND & MODEL OFFERED	UNIT PRICE	TOTAL PRICE
1	pH Meter (Handheld)	4 pcs			
	pH Range: 0.00 - 14.0 pH				
	Temperature Range: 0-55°C; 32-130 °F				
	Resolution : pH: 0.01 pH				
	Temperature 0.1 °C/°F				
	Accuracy : +/- 0.1 pH				
	Temperature Compensation : Automatic (ATC) to 25 C (ATC Range 1-50C)				
	Housing: Waterproof (submersible): floats				
	Calibration : Auto Calibration				
	Auto Shut-off: After 5 minutes				
	Probe: Glass Sensor and glass Reference tube				
	Safe Operating Temperature: 0-50 C; 32-122 F				
	Minumim Conductivity: For accuate and stable readings, the liquid to be tested ,must have a conductivity of atleast 10 µS (approx. 5ppm in TDS)				
	Display: LCD panel				
	Power source: 2X1.5V AAA (Alkaline batteries)				
2	Salinometer (Handheld)	4 pcs			
	Material : ABS				
	Item Name: Pen type Salinity Meter				
	Display: Large LCD Display with swittchable Backlight				
	Measuring Range: 0.00ppt-9.99ppt, 10.1ppt-50ppt				
	Resolution: 0.1ppt, 0.1 ppt				
	Accuracy: +/-3% FS +/- 1 digit				
	Repeatability: +/- 0.1ppt/.1ppt				
	Solvent Temperature Range: 0-60°C				
	Tempreature Accuracy: +/- 1° C				
	Working Temperature 0°C - 50°C				
	Working Humidity: 0%RH ~ 80%RH(Non-condensing)				
	Auto Power off: 5 Minutes without any operation				
	Power Supply: 4 * 1.5V LR44 Button Cell Battery (NOT Included)				
	BY LOT				
	NOTE: Please attach brochure and indicate days of delivery				

Accomplished by:

Supplier's Rppentative
(Print name and Signature)

Date Accomplished : _____

By the authority of the University President.

DR. DOLLY P. MAROMA
BAC Chairman

Canvassed by:

Name and Signature

BulSU-OP-PU-03F3
Revision: 1