## BIDS AND AWARDS COMMITTEE FOR GOODS, SUPPLIES, AND SERVICES

December 9, 2022

# Bid Bulletin No. 3 Modifications in the Bid Data Sheet, Schedule of Requirements, and Technical Specifications

This Bid Bulletin no. 3 is issued to modify or amend items in the Bidding Documents for project: "Procurement of Binocular Compound Microscope and Autoclave for San Rafael Campus (G-2023-01)". This shall form an integral part of the said Bidding Documents.

### I. SECTION III. BID DATA SHEET

ITB Clause	FROM				то			
5.3	For this purpose, contracts similar to the Project shall be:				For this purpose, contracts similar to the Project shall be:			
		imilar aborator <u>y</u>	Project: y Equipment	Clinical	a.	Similar Laboratory <u>Technical</u> <u>Equipment</u>	and	Clinica ment <u>c</u> Scientifi
	(3 th	Completed within the last three (3) years prior to the deadline for the submission and receipt of bids.		b.	Completed within the last three (3) years prior to the deadline for the submission and receipt of bids.			

### II. SECTION VI. SCHEDULE OF REQUIREMENTS

Item	FROM	ТО		
1	Binocular Compound Microscope and	Binocular Compound Microscope and		
	Autoclave	Autoclave		
	Within Fifteen (15) calendar days upon receipt of Notice to Proceed.	Within <u>Sixty (60) calendar days</u> upon receipt of Notice to Proceed.		

### III. SECTION VII. TECHNICAL SPECIFICATIONS

Item	FROM	ТО		
1	80 Units – BINOCULAR COMPOUND MICROSCOPE	80 Units – BINOCULAR COMPOUND MICROSCOPE		
	<ul> <li>Stage size: 140 x 140 mm</li> <li>Travel range X&amp;Y: 77 x 55 mm</li> <li>Illumination: Rechargeable LED illumination, 0.5W with intensity control</li> </ul>	<ul> <li>Stage size: 139-145 x 139-145 mm or higher</li> <li>Travel range X&amp;Y: 77 x 55 mm or higher</li> <li>Illumination: Rechargeable LED illumination, 0.5W minimum with intensity control</li> </ul>		
2	3 units - AUTOCLAVE	3 units - AUTOCLAVE		
	<ul> <li>Specifications:</li> <li>Sterilizing time = 7 minutes at 134°C</li> <li>With reproducible results</li> <li>Capacity: 16 L</li> <li>Inner dimension Ø 230 x 400 mm; 2 trays</li> </ul>	Specifications:  Sterilizing time = <u>at least</u> 7 minutes at 134°C  With reproducible results Deleted Capacity: 16 L ( <u>Maximum Capacity</u> ) Inner dimension Ø 230 x 400 mm; 2 trays Deleted		

#### Note:

- 1. Item, description or specifications <u>not mentioned above are without modifications</u> and <u>will remain as it is</u> in the bidding documents
- 2. Modified/additional entries are underlined, bold and italicized.

This Supplemental/ Bid Bulletin no. 3 is issued this  $\underline{9^{th}}$  day of December, 2022 for guidance and information of all concerned.

Dr. DOLLY P. MAROMA

Chairperson, BAC for Goods, Supplies & Services